

11-2020

A coded aperture microscope for X-ray fluorescence full-field imaging

D. P. Siddons

A. J. Kuczewski

A. K. Rumaiz

R. Tappero

M Idir

See next page for additional authors

Follow this and additional works at: https://digitalcommons.lsu.edu/physics_astronomy_pubs

 Part of the [Optics Commons](#)

Authors

D. P. Siddons, A. J. Kuczewski, A. K. Rumaiz, R. Tappero, M Idir, K. Nakhoda, J. Khanfri, V. Singh, E. R. Farquhar, M. Sullivan, D. Abel, D. J. Brady, and X. Yuan

A coded aperture microscope for X-ray fluorescence full-field imaging

D. P. Siddons,^a A. J. Kuczewski,^a A. K. Rumaiz,^a R. Tappero,^a M. Idir,^a
K. Nakhoda,^{a*} J. Khanfri,^b V. Singh,^c E. R. Farquhar,^d M. Sullivan,^d
D. Abel,^d D. J. Brady^e and X. Yuan^f

Received 25 June 2020
Accepted 7 September 2020

Edited by Y. Amemiya, University of Tokyo,
Japan

Keywords: coded aperture; X-ray microscope;
full-field; fluorescence.

^aNational Synchrotron Light Source II, Brookhaven National Laboratory, Upton, NY 11973, USA, ^bStony Brook University, Stony Brook, NY 11794, USA, ^cLouisiana State University, Baton Rouge, LA 70803, USA, ^dCase Western Reserve University, Cleveland, OH 44106, USA, ^eDuke University, Durham, NC 27708, USA, and ^fNokia Bell Labs, Murray Hill, NJ 07974, USA. *Correspondence e-mail: knakhoda@bnl.gov


The design and construction of an instrument for full-field imaging of the X-ray fluorescence emitted by a fully illuminated sample are presented. The aim is to produce an X-ray microscope with a few micrometers spatial resolution, which does not need to scan the sample. Since the fluorescence from a spatially inhomogeneous sample may contain many fluorescence lines, the optic which will provide the magnification of the emissions must be achromatic, *i.e.* its optical properties must be energy-independent. The only optics which fulfill this requirement in the X-ray regime are mirrors and pinholes. The throughput of a simple pinhole is very low, so the concept of coded apertures is an attractive extension which improves the throughput by having many pinholes, and retains the achromatic property. Modified uniformly redundant arrays (MURAs) with 10 μm openings and 50% open area have been fabricated using gold in a lithographic technique, fabricated on a 1 μm -thick silicon nitride membrane. The gold is 25 μm thick, offering good contrast up to 20 keV. The silicon nitride is transparent down into the soft X-ray region. MURAs with various orders, from 19 up to 73, as well as their respective negative (a mask where open and closed positions are inverted compared with the original mask), have been made. Having both signs of mask will reduce near-field artifacts and make it possible to correct for any lack of contrast.

1. Introduction

Current X-ray microprobe beamlines raster-scan the sample through a focused X-ray beam, and use an energy-resolving point detector to separate the contributions of the various elements at each point on the sample. Although this method has many advantages, it suffers from the need to mechanically move the sample through the beam, which is inevitably rather slow, particularly if the requirement is for 3D tomographic imaging, since this requires two translational scans for each of many rotational positions. An instrument capable of recording an image directly eliminates two of those dimensions, and can replace the rotation with a single translation for tomography (Takeuchi *et al.* 2009, 2010). A full-field microscope should collect fluorescence from all points on the sample simultaneously and project them onto an imaging detector. To image elemental contributions we need:

(i) An achromatic imaging optical element to magnify the fluorescence signal to match the desired resolution to the detector pixel size.

(ii) A detector which collects energy spectra from each detector pixel.



In this paper we demonstrate an achromatic imaging system, the coded aperture array. The hyperspectral imaging detector is also under development at Brookhaven National Laboratory, but is the subject of another work.

2. Coded apertures

2.1. Modified uniformly redundant arrays

The pinhole camera is the oldest and simplest imaging optic. It suffers from the inverse relationship between light collection efficiency and spatial resolution. The coded aperture array was developed as a way to break that relationship, by using arrays of pinholes in known locations. Even though the resulting image $I(x,y)$ is an overlay of many weak images, it is in principle possible to reconstitute the image mathematically. This became realistic as the era of digital computers arrived.

When the photons generated by an object $O(x,y)$ propagate through the coded aperture mask of binary function $M(x,y)$, the resulting image projected onto the detector is $I(x,y)$,

$$I(x, y) = O(x, y) \otimes M(x, y), \quad (1)$$

with \otimes denoting a convolution.

Various forms of coded aperture arrays have been demonstrated, with perhaps the most successful being the modified uniformly redundant array (MURA) proposed by Gottesman & Fenimore (1989). It has the advantage of having both high efficiency (50% open area) and a perfect system point spread function (SPSF). This means that the convolution of the MURA mask $M(x,y)$ with its anti-mask $H(x,y)$ is a perfect delta function,

$$M(x, y) \otimes H(x, y) = \delta. \quad (2)$$

A MURA mask $M(x,y)$ of order p , a prime number, and its anti-mask $H(x,y)$ are square lattices of size $p \times p$, with x and y varying from 0 to $p - 1$ and are defined as

$$M(x, y) = \begin{cases} 0, & \text{if } y = 0, \\ 1, & \text{if } x = 0, y \neq 0, \\ 1, & \text{if } x \text{ and } y \text{ are quadratic residues modulo } p, \\ 1, & \text{if neither } x \text{ or } y \text{ are quadratic residues modulo } p, \\ 0, & \text{otherwise.} \end{cases} \quad (3)$$

$$H(x, y) = \begin{cases} +1, & \text{if } x + y = 0, \\ +1, & \text{if } A(x, y) = 1 \quad (x + y \neq 0), \\ -1, & \text{if } A(x, y) = 0 \quad (x + y \neq 0). \end{cases} \quad (4)$$

The retrieved object is then obtained from the detected image on the detector as

$$O'(x, y) = I(x, y) \otimes H(x, y). \quad (5)$$

For this work, we chose to construct MURA masks with a unit pinhole of $10 \mu\text{m} \times 10 \mu\text{m}$ square, and designed arrays of order 19, 37 and 73. For a given pinhole size, the higher the order of the mask, the bigger the field of view of the system. Fig. 1(a) shows the pattern for order 73. It is usual to tile such




Figure 1 Pattern of an order 73 (a) single and (b) tiled MURA.

structures to increase its useful field of view. Shown in Fig. 1(b) is the layout of the typical 2×2 tiling that we sent for manufacturing.

2.2. Optic fabrication

Looking at the pattern of Fig. 1, it is obvious that the MURA mask is not self-supporting. Previous attempts to use such apertures have been forced to use self-supporting variants. Typically they were made by drilling or etching small holes in sheet metal such that no two holes touched (so-called NTHHT structures). Although these experiments have been proven successful (Kulow *et al.*, 2020; Haboub *et al.*, 2014), this involved a significant loss of solid angle, with the open area reducing to at best 25%. We wanted to achieve the theoretical maximum open area of about 50%, so needed a supported structure. A thin layer of silicon nitride is ideally suited for such a support, and this is what we implemented. The optics were fabricated using the LIGA technique (Guckel *et al.*, 1991), at the X-ray lithography facility at CAMD, in Baton Rouge, Louisiana, USA. We prepared lithography masks with all of the patterns listed above, plus other objects which might be of interest (single pinholes, slits and more). Silicon wafers were coated with a $1 \mu\text{m}$ -thick layer of silicon nitride. One side was patterned with squares matched to the array sizes, and the nitride etched away in those locations. An orientation-dependent silicon etch removed the silicon to leave a thin nitride membrane window which is X-ray transparent. On the other side of the wafer, a seed layer of gold was deposited, followed by a deep lithography in a thick photoresist. Following development of the resist, gold was electroplated into the cavities remaining, forming the thick gold aperture structures. Finally the photoresist was removed and the seed layer etched away, leaving free-standing structures. Fig. 2 shows the finished wafer, and Fig. 3 shows electron micrographs of one of the structures thus formed.

3. Experiments

3.1. Point source

Our first experiments used a microfocus X-ray generator (XRG) which was capable of providing a $10 \mu\text{m}$ focal spot from a molybdenum anode tube. We used the order 73 MURA to image that focal spot onto a Medipix detector. This detector




Figure 2
The finished silicon wafer prepared using the LIGA technique at the X-ray lithography facility at CAMD. It holds MURA of order 19, 37 and 73 as well as other objects such as single pinholes and slits.




Figure 3
Electron micrographs of the MURA coded aperture with 10 μm holes manufactured using the LIGA technique.

has a 256×256 array of photon-counting pixels on a 55 μm pitch. It does not provide energy resolution, but can accept a high photon flux. The resulting image is shown in Fig. 4(a). Since the source is essentially a point source, the recorded image is a smeared shadow of the mask. It was easy to perform the reconstruction using total variation regularized least-




Figure 4
(a) Image of an XRG point source through an order 73 MURA on a 256×256 pixels Medipix detector. (b) Retrieved point source from the shadow of the XRG source through the order 73 MURA on the Medipix detector using total variation least-squares deconvolution.

squares deconvolution (Chan *et al.*, 2011), and the result is shown in Fig. 4(b).

3.2. Extended source

We used the XFP beamline at NSLS-II (Asuru *et al.*, 2019) to demonstrate that the image processing method would also work on extended and fluorescing samples. XFP can operate in a pink beam mode with a partially focused spot of roughly $2 \text{ mm} \times 2 \text{ mm}$. The optics of the beamline were optimized for low-energy photons, mainly below 20 keV, with the majority of the beam power below 10 keV. We added a 1 mm-thick aluminium filter upstream of our setup to reduce the thermal load this beam would place on any sample. Fig. 5 shows the resultant calculated spectrum.

We chose a gold wire mesh as our sample. The mesh is nominally 25 μm wires woven into a 300 μm -pitch square mesh (Fig. 6). It was placed across a rectangular frame, such that the incident beam only illuminated the mesh and air. The distance d_1 between the object and the mask was 38 mm and




Figure 5
Calculated spectrum of the XFP pink beam with 1 mm-thick aluminium filter.




Figure 6
Optical micrograph of the gold wire mesh with 25 μm wires woven into a 300 μm -pitch square mesh.




Figure 7
 (a) Raw data on the 256×256 pixels Medipix detector from XFP fluorescence measurement of the gold wire mesh. The beam measured $200 \mu\text{m} \times 200 \mu\text{m}$ and the mesh was oriented with its wires horizontal and vertical. (b) Retrieved image of the gold wire mesh obtained from using total variation least-squares deconvolution of the detector's raw data.

the distance d_2 between the mask and the detector was 571 mm. We used the same Medipix detector as the one described in Section 3.1. Given the geometry of the mask and the setup, the theoretical resolution r is calculated to be $14.3 \mu\text{m}$ as

$$r = h \left(1 + \frac{d_1}{d_2} \right) + Pd \frac{d_1}{d_2}, \quad (6)$$

with h the size of the mask pinhole, $10 \mu\text{m}$ in our case, and Pd the size of the detector pixel which is $55 \mu\text{m}$.

The first image had the mesh oriented with its wires horizontal and vertical. It was illuminated with a small beam, approximately $200 \mu\text{m} \times 200 \mu\text{m}$. The raw data and reconstructed image are shown in Fig. 7. We observed approximately $22.4 \mu\text{m}$ wire width, which suggests that the actual resolution of the setup is less than $25 \mu\text{m}$. The discrepancy in the wire width could easily come from uncertainty in the magnification which will be properly characterized in future measurements.

We then oriented the mesh so that the wires were at roughly 45° to the horizontal, and enlarged the illuminating beam to $600 \mu\text{m} \times 400 \mu\text{m}$. The raw data and reconstruction are shown in Fig. 8. The contrast is reduced compared with the image with the smaller beam. We believe that this is due to the increased parasitic scattering from the larger beam, adding a




Figure 8
 (a) Raw data on the 256×256 pixels Medipix detector from XFP fluorescence measurement of the gold wire mesh. The beam measured $600 \mu\text{m} \times 400 \mu\text{m}$ and the mesh was oriented at 45° . (b) Retrieved image of the gold wire mesh obtained from using total variation least-squares deconvolution of the detector's raw data.

uniform background to the data. This is probably a consequence of our decision to operate close to backscattering, rather than at 90° to the incident beam as is usually done for fluorescence measurements. Nevertheless, the image can be reconstructed successfully. Similarly to the point source, the images of the extended source [Figs. 7(b) and 8(b)] were reconstructed using total variation regularized least-squares deconvolution.

4. Conclusions

Using the LIGA technique, we fabricated a MURA of order 73 with $10 \mu\text{m}$ holes and 50% open area, providing the maximum possible throughput for such a system. Using this optic we could reconstruct part of a gold wire mesh of $25 \mu\text{m}$ wires on a $300 \mu\text{m}$ pitch. This suggests that we reach a resolution better than $25 \mu\text{m}$. The theoretical resolution for our setup is $14.3 \mu\text{m}$. In our future work we will carefully characterize the magnification of the system and upgrade our image processing methods as suggested by Kulow *et al.* (2020). We will also design and fabricate apertures which match more closely the needs of the intended research, and develop a hybrid pixel detector in which each pixel is a spectrometer capable of handling a high rate and providing high-quality spectra.

Funding information

The full-field fluorescence imaging (FFFI) development presented in this publication was supported by the US Department of Energy, Office of Science, Biological and Environmental Research under Award FWP PS018. Parts of this research utilized the XFP (17-BM) beamline. Funding for development of XFP was provided by the National Science Foundation, Division of Biological Infrastructure (grant No. 1228549). Operations support of XFP was via the National Institutes of Health (grant No. P30-EB-009998). Use of the National Synchrotron Light Source II, a US Department of Energy (DOE) Office of Science User Facility operated for the DOE Office of Science by Brookhaven National Laboratory, was supported under Contract No. DE-SC0012704.

References

Asuru, A., Farquhar, E. R., Sullivan, M., Abel, D., Toomey, J., Chance, M. R. & Bohon, J. (2019). *J. Synchrotron Rad.* **26**, 1388–1399.
 Chan, S. H., Khoshabeh, R., Gibson, K. B., Gill, P. E. & Nguyen, T. Q. (2011). *IEEE Trans. Image Process.* **20**, 3097–3111.
 Gottesman, S. R. & Fenimore, E. E. (1989). *Appl. Opt.* **28**, 4344–4352.
 Guckel, K., Skrobis, K. J., Christenson, T. R., Klein, J., Han, S., Choi, B. & Lovell, E. G. (1991). *Proceedings of the IEEE International Conference on Micro Electro Mechanical Systems*, 30 January–2 February 1991, Nara, Japan, pp. 74–79.
 Haboub, A., MacDowell, A. A., Marchesini, S. & Parkinson, D. Y. (2014). *Rev. Sci. Instrum.* **85**, 063704.
 Kulow, A., Buzanich, A. G., Reinholz, U., Strelt, C. & Radtke, M. (2020). *J. Anal. At. Spectrom.* **35**, 347–356.
 Takeuchi, A., Terada, Y., Suzuki, Y., Uesugi, K. & Aoki, S. (2009). *J. Synchrotron Rad.* **16**, 616–621.
 Takeuchi, A., Terada, Y., Uesugi, K. & Suzuki, Y. (2010). *Nucl. Instrum. Methods Phys. Res. A*, **616**, 261–265.